
Midmark M11 - v1.0.0
Total Cycles: 4570

Name: RITTER 1
Sterilizer ID: V2178198
05/22/2024 07:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 70.2 F 0.2
4:00 80.6 F 1.0
6:00 108.8 F 3.1
8:00 137.7 F 5.7
10:00 162.0 F 9.0
12:00 181.8 F 12.7
14:00 200.1 F 3.5
16:00 212.3 F 6.8
18:00 222.7 F 10.3
20:00 230.2 F 8.7
22:00 235.7 F 10.1
24:00 236.0 F 8.6
26:00 245.4 F 13.1
28:00 256.2 F 19.0
30:00 266.3 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.6 F 29.1
2:00 271.7 F 28.8
3:00 272.4 F 29.1
4:00 271.7 F 28.7

Min 271.6 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 198.9 F 0.0
10:00 204.1 F 0.0
15:00 209.0 F 0.0
20:00 214.4 F 0.0
25:00 217.2 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:43
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:09:50
CYCLE COMPLETE

05/22/2024 09:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____