
Midmark M11 - v1.0.0
Total Cycles: 2166

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2022 08:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.9 F 0.0
2:00 175.9 F 0.6
4:00 175.2 F 4.5
6:00 193.9 F 8.9
8:00 222.9 F 9.3
10:00 224.5 F 8.5
12:00 237.4 F 13.0
14:00 238.5 F 10.6
16:00 235.6 F 8.2
18:00 248.6 F 14.9
20:00 260.5 F 21.3
22:00 271.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 272.7 F 29.3
2:00 271.9 F 29.0
3:00 272.6 F 29.3
4:00 271.8 F 28.9

Min 271.8 F 28.7
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 230.5 F 0.0
10:00 220.9 F 0.0
15:00 227.1 F 0.0
20:00 230.4 F 0.0
25:00 232.3 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 23:31
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:30
CYCLE COMPLETE

07/25/2022 09:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____