
Midmark M11 - v1.0.5
Total Cycles: 2113

Name: RITTER2
Sterilizer ID: V2313228
03/02/2023 09:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.1 F 0.0
2:00 131.2 F 1.2
4:00 145.9 F 4.1
6:00 174.5 F 8.3
8:00 198.8 F 12.3
10:00 220.2 F 7.1
12:00 232.5 F 12.7
14:00 236.3 F 10.2
16:00 235.1 F 9.1
18:00 250.6 F 16.3
20:00 263.0 F 23.5
22:00 271.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.2 F 29.5
2:00 272.5 F 30.0
3:00 271.5 F 29.2
4:00 272.5 F 29.8

Min 271.5 F 29.2
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.0 F 0.0
5:00 210.4 F 0.0
10:00 233.7 F 0.0
15:00 222.9 F 0.0
20:00 222.7 F 0.0
25:00 233.9 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 22:03
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:19
CYCLE COMPLETE

03/02/2023 10:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____