

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000954  
13:02 25/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:05  
PRESSURIZING 4:10  
STERILIZING 5:19  
136.0 C 227kPa 5:19  
136.2 C 231kPa 5:49  
136.3 C 232kPa 6:19  
136.3 C 231kPa 6:49  
136.4 C 230kPa 7:19  
136.4 C 233kPa 7:49  
136.3 C 231kPa 8:19  
136.3 C 232kPa 8:49  
136.3 C 232kPa 8:49  
VENTING 8:50  
AIR DRYING 9:43  
CYCLE COMPLETE 24:43  
-----

Digital Signature #  
16D2E5046867DAC2  
-----