
Midmark M11 - v1.0.5
Total Cycles: 1802

Name: RITTER2
Sterilizer ID: V2313228
12/05/2022 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.7 F 0.0
2:00 180.8 F 3.4
4:00 197.6 F 8.1
6:00 215.9 F 10.2
8:00 230.1 F 9.7
10:00 233.7 F 8.7
12:00 242.6 F 11.9
14:00 246.0 F 13.6
16:00 258.8 F 20.6
18:00 270.1 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.7 F 29.6
2:00 271.6 F 28.9
3:00 272.0 F 29.5
4:00 272.3 F 29.4

Min 271.2 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.2 F 0.0
5:00 226.8 F 0.0
10:00 236.5 F 0.0
15:00 228.2 F 0.0
20:00 226.9 F 0.0
25:00 227.9 F 0.0
30:00 226.4 F 0.0

Drying Complete

FILLING: 1:54
HEATING: 18:56
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:59
CYCLE COMPLETE

12/05/2022 02:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____