

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2909

Name: RITTER2  
Sterilizer ID: V2313228  
09/20/2023 11:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.3 F 0.0  
2:00 166.4 F 2.6  
4:00 185.8 F 6.5  
6:00 205.2 F 10.9  
8:00 213.8 F 5.6  
10:00 226.2 F 9.5  
12:00 237.5 F 10.2  
14:00 241.3 F 11.8  
16:00 238.9 F 10.3  
18:00 250.9 F 16.3  
20:00 262.3 F 22.5  
22:00 271.9 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 272.7 F 29.6  
2:00 272.2 F 30.0  
3:00 271.4 F 29.3  
4:00 272.7 F 29.9

Min 271.2 F 29.2  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.3 F 0.0  
5:00 233.9 F 0.0  
10:00 190.4 F 0.0  
15:00 190.5 F 0.0  
20:00 183.5 F 0.0  
25:00 181.1 F 0.0  
30:00 182.7 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 23:26  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:02:27  
CYCLE COMPLETE

09/20/2023 12:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_