

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3045

Name: RITTER 1  
Sterilizer ID: V2178198  
03/28/2023 11:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 142.5 F 0.0  
2:00 149.1 F 1.0  
4:00 169.0 F 4.7  
6:00 190.7 F 8.9  
8:00 212.1 F 13.3  
10:00 219.8 F 7.4  
12:00 234.2 F 12.4  
14:00 234.8 F 9.2  
16:00 239.2 F 9.6  
18:00 246.6 F 13.6  
20:00 258.8 F 20.2  
22:00 269.8 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.3 F 28.9  
2:00 271.8 F 29.1  
3:00 272.0 F 28.7  
4:00 272.1 F 29.2

Min 271.4 F 28.4  
Max 272.4 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.7 F 0.0  
5:00 216.0 F 0.0  
10:00 217.4 F 0.0  
15:00 231.9 F 0.0  
20:00 228.1 F 0.0  
25:00 230.4 F 0.0  
30:00 230.8 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 22:55  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:01:04  
CYCLE COMPLETE

03/28/2023 12:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_