
Midmark M11 - v1.0.5
Total Cycles: 4187

Name: RITTER2
Sterilizer ID: V2313228
12/05/2024 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 73.8 F 0.0
4:00 78.4 F 0.0
6:00 98.3 F 0.1
8:00 149.0 F 0.2
10:00 177.0 F 0.5
12:00 196.7 F 2.5
14:00 209.8 F 5.4
16:00 219.1 F 8.6
18:00 233.5 F 12.2
20:00 229.3 F 7.8
22:00 240.2 F 12.4
24:00 237.9 F 10.3
26:00 237.5 F 9.0
28:00 248.7 F 15.1
30:00 259.5 F 21.2
32:00 269.6 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.6
2:00 271.6 F 28.9
3:00 272.7 F 29.8
4:00 271.9 F 29.1

Min 271.4 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 215.3 F 0.0
10:00 190.0 F 0.0
15:00 185.0 F 0.0
20:00 183.7 F 0.0
25:00 184.7 F 0.0
30:00 189.0 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 33:03
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:11:33
CYCLE COMPLETE

12/05/2024 11:14 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____