
Midmark M11 - v1.0.5
Total Cycles: 3309

Name: RITTER2
Sterilizer ID: V2313228
01/19/2024 11:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.4 F 0.0
2:00 149.0 F 1.4
4:00 160.1 F 4.6
6:00 187.2 F 8.7
8:00 204.3 F 13.6
10:00 221.0 F 7.7
12:00 231.0 F 12.7
14:00 235.1 F 9.8
16:00 237.1 F 9.3
18:00 247.7 F 14.6
20:00 259.6 F 21.3
22:00 270.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.5 F 29.6
2:00 271.3 F 29.0
3:00 272.6 F 29.8
4:00 272.0 F 29.3

Min 271.2 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 235.0 F 0.0
10:00 224.9 F 0.0
15:00 211.8 F 0.0
20:00 217.5 F 0.0
25:00 219.7 F 0.0
30:00 215.8 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:50
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:01:20
CYCLE COMPLETE

01/19/2024 12:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____