
Midmark M11 - v1.0.5
Total Cycles: 346

Name: RITTER2
Sterilizer ID: V2313228
09/09/2021 10:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.7 F 0.0
2:00 179.0 F 3.1
4:00 202.6 F 8.1
6:00 219.2 F 10.5
8:00 232.5 F 9.7
10:00 235.5 F 9.2
12:00 243.1 F 12.1
14:00 246.0 F 13.7
16:00 259.1 F 20.7
18:00 270.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 271.9 F 28.9
2:00 272.3 F 29.2
3:00 272.6 F 29.6
4:00 271.8 F 29.1

Min 271.6 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 237.0 F 0.0
10:00 211.6 F 0.0
15:00 216.3 F 0.0
20:00 208.3 F 0.0
25:00 210.9 F 0.0
30:00 203.4 F 0.0
Drying Complete

FILLING: 2:41
HEATING: 18:44
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

09/09/2021 12:06 PM
ACCEPTED: X
REJECTED:
Operator: U02