
Midmark M11 - v1.0.5
Total Cycles: 3827

Name: RITTER2
Sterilizer ID: V2313228
08/01/2024 10:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.7 F 0.0
2:00 121.3 F 0.3
4:00 131.7 F 2.2
6:00 158.1 F 5.2
8:00 180.5 F 8.9
10:00 198.4 F 12.7
12:00 212.0 F 5.3
14:00 222.9 F 9.0
16:00 233.9 F 13.3
18:00 234.9 F 9.9
20:00 237.9 F 9.7
22:00 245.4 F 13.5
24:00 256.4 F 19.4
26:00 266.7 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.7 F 29.7
2:00 271.7 F 29.2
3:00 272.8 F 29.7
4:00 271.6 F 29.0

Min 271.6 F 29.0
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 214.3 F 0.0
10:00 223.9 F 0.0
15:00 221.7 F 0.0
20:00 217.3 F 0.0
25:00 215.4 F 0.0
30:00 214.4 F 0.0

Drying Complete

FILLING: 0:54
HEATING: 27:37
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:06:09
CYCLE COMPLETE

08/01/2024 11:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____