

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1373

Name: RITTER2  
Sterilizer ID: V2313228  
08/08/2022 09:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.8 F 0.0  
2:00 143.3 F 0.0  
4:00 151.4 F 0.0  
6:00 185.3 F 0.7  
8:00 205.0 F 4.9  
10:00 218.4 F 9.5  
12:00 230.1 F 9.5  
14:00 236.3 F 10.9  
16:00 234.7 F 8.8  
18:00 239.4 F 10.1  
20:00 248.1 F 14.7  
22:00 260.3 F 21.5  
24:00 271.2 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.6 F 29.4  
2:00 271.5 F 28.9  
3:00 272.7 F 29.4  
4:00 271.5 F 28.7

Min 271.4 F 28.7  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.5 F 0.0  
5:00 207.8 F 0.0  
10:00 229.2 F 0.0  
15:00 227.0 F 0.0  
20:00 221.3 F 0.0  
25:00 217.9 F 0.0  
30:00 218.3 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 24:40  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:02:42  
CYCLE COMPLETE

08/08/2022 10:12 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_