
Midmark M11 - v1.0.0
Total Cycles: 3543

Name: RITTER 1
Sterilizer ID: V2178198
07/18/2023 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.4 F 0.0
2:00 201.1 F 3.2
4:00 203.9 F 8.1
6:00 224.7 F 10.7
8:00 230.3 F 9.3
10:00 235.6 F 9.4
12:00 242.3 F 12.6
14:00 242.0 F 11.4
16:00 255.2 F 18.1
18:00 266.7 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.7 F 29.1
2:00 271.7 F 28.4
3:00 272.2 F 29.0
4:00 272.4 F 28.9

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 221.0 F 0.0
10:00 232.2 F 0.0
15:00 230.2 F 0.0
20:00 224.4 F 0.0
25:00 220.4 F 0.0
30:00 217.8 F 0.0

Drying Complete

FILLING: 1:50
HEATING: 19:32
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

07/18/2023 12:16 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____