
Midmark M11 - v1.0.0
Total Cycles: 4001

Name: RITTER 1
Sterilizer ID: V2178198
11/09/2023 12:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.2 F 0.0
2:00 163.2 F 2.8
4:00 171.9 F 6.5
6:00 193.7 F 10.6
8:00 214.7 F 4.7
10:00 221.3 F 8.0
12:00 231.2 F 12.6
14:00 233.6 F 9.2
16:00 239.3 F 10.4
18:00 245.6 F 13.3
20:00 257.0 F 19.3
22:00 267.3 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.6 F 29.4
2:00 272.1 F 28.9
3:00 271.2 F 28.6
4:00 272.2 F 29.4

Min 271.2 F 28.5
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.2 F 0.0
5:00 222.8 F 0.0
10:00 241.9 F 0.0
15:00 230.0 F 0.0
20:00 222.9 F 0.0
25:00 220.4 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 23:31
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:02:25
CYCLE COMPLETE

11/09/2023 01:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____