
Midmark M11 - v1.0.5
Total Cycles: 1898

Name: RITTER2
Sterilizer ID: V2313228
01/05/2023 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.3 F 0.0
2:00 172.0 F 2.5
4:00 170.8 F 5.9
6:00 191.1 F 10.5
8:00 212.7 F 4.6
10:00 226.0 F 9.7
12:00 232.9 F 8.9
14:00 241.3 F 12.6
16:00 240.9 F 11.4
18:00 253.5 F 17.9
20:00 265.1 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.5 F 29.6
2:00 271.8 F 29.6
3:00 272.3 F 29.6
4:00 271.6 F 29.3

Min 271.5 F 29.1
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 235.2 F 0.0
10:00 228.3 F 0.0
15:00 220.7 F 0.0
20:00 217.4 F 0.0
25:00 218.1 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 21:46
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

01/05/2023 11:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____