
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001671
16:13 03/01/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 2:22
STERILIZING 3:21
136.0 C 220kPa 3:21
136.3 C 222kPa 3:51
136.4 C 225kPa 4:21
136.2 C 223kPa 4:51
136.3 C 223kPa 5:21
136.3 C 225kPa 5:51
136.4 C 226kPa 6:21
136.4 C 226kPa 6:51
136.4 C 226kPa 6:51
VENTING 6:52
AIR DRYING 7:42
CYCLE COMPLETE 7:57

Digital Signature #
6C02B9A77B321549
