
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001848
10:57 15/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:13
STERILIZING 4:18
136.4 C 227kPa 4:18
136.2 C 224kPa 4:48
136.3 C 225kPa 5:18
136.4 C 223kPa 5:48
136.4 C 226kPa 6:18
136.3 C 225kPa 6:48
136.3 C 224kPa 7:18
136.4 C 225kPa 7:48
136.4 C 225kPa 7:48
VENTING 7:48
AIR DRYING 8:46
CYCLE COMPLETE 9:20

Digital Signature #
26DF441D5D1DA84F
