
Midmark M11 - v1.0.5
Total Cycles: 3361

Name: RITTER2
Sterilizer ID: V2313228
02/06/2024 03:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.1 F 0.0
2:00 125.7 F 0.7
4:00 152.2 F 3.4
6:00 179.2 F 7.2
8:00 198.8 F 11.4
10:00 212.3 F 4.8
12:00 224.7 F 9.1
14:00 233.8 F 11.4
16:00 238.3 F 11.1
18:00 238.0 F 9.7
20:00 251.0 F 16.2
22:00 262.4 F 22.9
24:00 272.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 272.4 F 29.7
2:00 272.7 F 29.3
3:00 271.7 F 29.1
4:00 272.8 F 29.9

Min 271.6 F 28.8
Max 273.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.6 F 0.0
5:00 212.1 F 0.0
10:00 207.0 F 0.0
15:00 211.6 F 0.0
20:00 213.7 F 0.0
25:00 216.7 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:12
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:02:40
CYCLE COMPLETE

02/06/2024 04:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____