
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003248
13:52 01/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:19
STERILIZING 4:23
136.4 C 222kPa 4:23
136.4 C 225kPa 4:53
136.0 C 221kPa 5:23
136.4 C 224kPa 5:53
136.5 C 224kPa 6:23
136.4 C 225kPa 6:53
136.4 C 225kPa 7:23
136.4 C 222kPa 7:53
136.4 C 222kPa 7:53
VENTING 7:53
AIR DRYING 8:49
CYCLE COMPLETE 9:03

Digital Signature #
05490DF07B5CB6E2
