
Midmark M11 - v1.0.5
Total Cycles: 3152

Name: RITTER2
Sterilizer ID: V2313228
11/21/2023 02:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.2 F 0.0
2:00 106.1 F 0.4
4:00 133.4 F 2.4
6:00 163.0 F 6.1
8:00 187.7 F 10.7
10:00 206.3 F 3.8
12:00 220.9 F 7.7
14:00 232.9 F 12.3
16:00 234.7 F 9.2
18:00 239.1 F 9.9
20:00 247.3 F 14.1
22:00 260.0 F 21.1
24:00 271.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.2 F 28.9
2:00 272.8 F 29.4
3:00 271.9 F 29.1
4:00 272.2 F 28.9

Min 271.6 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.6 F 0.0
5:00 200.7 F 0.0
10:00 210.7 F 0.0
15:00 211.1 F 0.0
20:00 209.1 F 0.0
25:00 209.3 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 24:33
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:59
CYCLE COMPLETE

11/21/2023 03:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____