
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000679
15:14 14/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:17
PRESSURIZING 1:23
STERILIZING 2:21
136.0 C 229kPa 2:21
136.2 C 232kPa 2:51
136.4 C 233kPa 3:21
136.4 C 233kPa 3:51
136.4 C 233kPa 4:21
136.0 C 228kPa 4:51
136.3 C 229kPa 5:21
136.1 C 232kPa 5:51
136.1 C 232kPa 5:51
VENTING 5:51
AIR DRYING 6:42
CYCLE COMPLETE 6:44

Digital Signature #
DBC479AB700EAB2A
