
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006140
14:10 16/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:36
STERILIZING 3:23
136.0 C 220kPa 3:23
135.8 C 221kPa 3:53
136.4 C 222kPa 4:23
136.1 C 223kPa 4:53
136.3 C 224kPa 5:23
136.3 C 223kPa 5:53
136.4 C 223kPa 6:23
136.4 C 224kPa 6:53
136.4 C 224kPa 6:53
VENTING 6:53
AIR DRYING 7:46
CYCLE COMPLETE 10:22

Digital Signature #
56DF597F68988578
