
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007199
6:55 15/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:13
STERILIZING 4:17
136.0 C 220kPa 4:17
136.2 C 220kPa 4:47
136.3 C 221kPa 5:17
136.4 C 223kPa 5:47
136.4 C 221kPa 6:17
136.2 C 220kPa 6:47
136.2 C 218kPa 7:17
136.3 C 222kPa 7:47
136.3 C 222kPa 7:47
VENTING 7:47
AIR DRYING 8:35
CYCLE COMPLETE 13:59

Digital Signature #
2BC468175827687E
