
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007264
8:18 25/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 2:59
STERILIZING 4:05
136.4 C 223kPa 4:05
136.0 C 222kPa 4:35
136.1 C 221kPa 5:05
136.4 C 224kPa 5:35
136.1 C 223kPa 6:05
136.4 C 222kPa 6:35
136.3 C 222kPa 7:05
136.4 C 221kPa 7:35
136.4 C 221kPa 7:35
VENTING 7:35
AIR DRYING 8:32
CYCLE COMPLETE 8:37

Digital Signature #
F338DCF46E2759E5
