
Midmark M11 - v1.0.0
Total Cycles: 945

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2021 09:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.1 F 0.0
2:00 110.5 F 0.0
4:00 138.1 F 0.1
6:00 159.6 F 1.0
8:00 182.5 F 3.6
10:00 195.1 F 6.3
12:00 207.8 F 10.0
14:00 219.5 F 8.4
16:00 226.3 F 7.8
18:00 236.9 F 12.0
20:00 235.2 F 8.7
22:00 244.1 F 12.2
24:00 244.5 F 12.3
26:00 255.7 F 18.1
28:00 265.8 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.5 F 28.8
2:00 271.7 F 28.1
3:00 272.0 F 28.6
4:00 272.3 F 28.6

Min 271.4 F 28.0
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.0 F 0.0
5:00 216.1 F 0.0
10:00 214.0 F 0.0
15:00 198.3 F 0.0
20:00 199.1 F 0.0
25:00 180.9 F 0.0
30:00 182.4 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 29:51
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:08:06
CYCLE COMPLETE

06/28/2021 10:12 AM
ACCEPTED: X
REJECTED:
Operator: U01