
Midmark M11 - v1.0.5
Total Cycles: 218

Name: RITTER2
Sterilizer ID: V2313228
07/15/2021 02:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.8 F 0.0
2:00 138.2 F 1.9
4:00 161.7 F 5.2
6:00 187.1 F 9.6
8:00 205.4 F 7.3
10:00 224.4 F 7.4
12:00 236.3 F 12.5
14:00 237.5 F 10.3
16:00 236.4 F 8.9
18:00 249.8 F 15.5
20:00 261.3 F 22.1
22:00 271.9 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.5 F 29.4
2:00 271.4 F 28.9
3:00 272.7 F 29.4
4:00 271.5 F 28.7

Min 271.3 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 230.7 F 0.0
10:00 230.7 F 0.0
15:00 225.8 F 0.0
Dry Cycle Aborted

FILLING: 1:08
HEATING: 22:33
STERILIZING: 4:00
VENTING: 3:12
DRYING: 18:40
TOTAL CYCLE: 00:49:34

07/15/2021 03:38 PM
ACCEPTED: X
REJECTED:
Operator: U02