

-----  
Midmark M11 - v1.0.5  
Total Cycles: 872

Name: RITTER2  
Sterilizer ID: V2313228  
03/10/2022 12:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.5 F 0.0  
2:00 160.7 F 3.1  
4:00 184.4 F 7.5  
6:00 205.3 F 12.9  
8:00 227.5 F 7.9  
10:00 239.4 F 13.3  
12:00 239.2 F 11.1  
14:00 237.7 F 9.8  
16:00 251.9 F 16.6  
18:00 263.7 F 23.5  
20:00 271.4 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.6 F 29.4  
2:00 271.3 F 29.1  
3:00 272.9 F 29.5  
4:00 271.2 F 28.8

Min 271.1 F 28.8  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.9 F 0.0  
5:00 234.5 F 0.0  
10:00 221.7 F 0.0  
15:00 204.4 F 0.0  
20:00 187.0 F 0.0  
25:00 222.2 F 0.0  
30:00 175.2 F 0.0  
Drying Complete

FILLING: 2:04  
HEATING: 20:06  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:18  
CYCLE COMPLETE

03/10/2022 01:25 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02