
Midmark M11 - v1.0.5
Total Cycles: 1080

Name: RITTER2
Sterilizer ID: V2313228
05/09/2022 06:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.4 F 0.0
2:00 79.0 F 0.0
4:00 85.5 F 0.0
6:00 107.4 F 0.0
8:00 170.0 F 0.2
10:00 197.8 F 2.9
12:00 198.6 F 6.5
14:00 212.0 F 10.1
16:00 221.3 F 9.1
18:00 232.8 F 8.8
20:00 241.1 F 13.4
22:00 239.0 F 10.8
24:00 236.1 F 8.9
26:00 250.4 F 15.5
28:00 261.6 F 22.0
30:00 272.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.5
1:00 272.6 F 29.3
2:00 271.6 F 28.5
3:00 271.8 F 29.2
4:00 272.7 F 29.4

Min 270.9 F 28.3
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.1 F 0.0
5:00 220.9 F 0.0
10:00 185.1 F 0.0
15:00 174.2 F 0.0
20:00 185.8 F 0.0
25:00 195.1 F 0.0
30:00 166.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:35
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:08:28
CYCLE COMPLETE

05/09/2022 08:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____