

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002279  
11:26 18/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:35  
STERILIZING 3:39  
136.0 C 222kPa 3:39  
136.4 C 225kPa 4:09  
136.1 C 223kPa 4:39  
136.1 C 223kPa 5:09  
136.4 C 226kPa 5:39  
136.2 C 223kPa 6:09  
136.4 C 226kPa 6:39  
136.4 C 226kPa 7:09  
136.4 C 226kPa 7:09  
VENTING 7:09  
AIR DRYING 8:02  
CYCLE COMPLETE 24:53  
-----

Digital Signature #  
8120B4CCEF7449E8  
-----