
Midmark M11 - v1.0.5
Total Cycles: 1886

Name: RITTER2
Sterilizer ID: V2313228
12/28/2022 12:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.5 F 0.0
2:00 200.1 F 3.6
4:00 200.1 F 9.2
6:00 221.7 F 7.4
8:00 233.7 F 11.1
10:00 234.6 F 9.1
12:00 237.6 F 9.4
14:00 249.7 F 15.5
16:00 262.2 F 22.8
18:00 272.5 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 272.1 F 29.8
2:00 271.8 F 29.1
3:00 271.5 F 29.4
4:00 272.1 F 29.5

Min 271.0 F 28.8
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 225.0 F 0.0
10:00 238.2 F 0.0
15:00 232.6 F 0.0
20:00 224.0 F 0.0
25:00 224.2 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 18:07
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:13
CYCLE COMPLETE

12/28/2022 01:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____