
Midmark M11 - v1.0.0
Total Cycles: 5238

Name: RITTER 1
Sterilizer ID: V2178198
01/13/2025 10:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 69.1 F 0.0
4:00 82.1 F 0.0
6:00 116.2 F 0.1
8:00 142.4 F 0.2
10:00 166.1 F 0.2
12:00 186.0 F 0.5
14:00 203.2 F 1.6
16:00 218.3 F 4.7
18:00 224.8 F 7.6
20:00 233.2 F 10.9
22:00 234.3 F 8.1
24:00 240.5 F 11.3
26:00 235.9 F 8.7
28:00 242.1 F 10.8
30:00 246.3 F 13.2
32:00 256.8 F 18.8
34:00 266.3 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.2 F 28.6
2:00 272.7 F 28.8
3:00 272.2 F 28.4
4:00 271.7 F 28.0

Min 271.6 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 189.0 F 0.0
10:00 222.5 F 0.0
15:00 221.4 F 0.0
20:00 214.8 F 0.0
25:00 211.9 F 0.0
30:00 213.1 F 0.0
Drying Complete

FILLING: 0:42
HEATING: 35:53
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:14:28
CYCLE COMPLETE

01/13/2025 11:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____