

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2501

Name: RITTER2  
Sterilizer ID: V2313228  
06/12/2023 01:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.8 F 0.0  
2:00 171.7 F 3.1  
4:00 180.4 F 6.7  
6:00 199.7 F 10.9  
8:00 217.6 F 5.1  
10:00 221.6 F 9.5  
12:00 232.0 F 10.3  
14:00 239.1 F 11.7  
16:00 237.8 F 10.2  
18:00 250.5 F 16.1  
20:00 261.0 F 22.2  
22:00 270.7 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.2  
1:00 271.5 F 29.7  
2:00 273.3 F 30.2  
3:00 271.9 F 30.0  
4:00 272.2 F 29.8

Min 271.0 F 29.1  
Max 273.3 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.5 F 0.0  
5:00 233.7 F 0.0  
10:00 219.9 F 0.0  
15:00 219.6 F 0.0  
20:00 220.1 F 0.0  
25:00 229.3 F 0.0  
30:00 226.2 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 22:45  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:01:41  
CYCLE COMPLETE

06/12/2023 02:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_