

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4403

Name: RITTER2  
Sterilizer ID: V2313228  
03/05/2025 03:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.9 F 0.0  
2:00 141.5 F 1.4  
4:00 152.5 F 4.3  
6:00 178.9 F 8.0  
8:00 198.0 F 12.0  
10:00 217.2 F 6.4  
12:00 225.3 F 11.0  
14:00 229.0 F 7.1  
16:00 240.2 F 12.6  
18:00 239.3 F 10.9  
20:00 250.1 F 16.9  
22:00 261.1 F 23.1  
24:00 270.9 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 30.4  
1:00 273.0 F 31.0  
2:00 271.8 F 30.3  
3:00 271.9 F 31.0  
4:00 273.2 F 30.8

Min 271.3 F 30.2  
Max 273.4 F 31.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.8 F 0.0  
5:00 200.7 F 0.0  
10:00 210.8 F 0.0  
15:00 217.6 F 0.0  
20:00 218.8 F 0.0  
25:00 217.6 F 0.0  
30:00 216.9 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 24:48  
STERILIZING: 4:00  
VENTING: 3:55  
DRYING: 30:00  
TOTAL CYCLE: 01:03:49  
CYCLE COMPLETE

03/05/2025 04:49 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_