
Midmark M11 - v1.0.5
Total Cycles: 3956

Name: RITTER2
Sterilizer ID: V2313228
09/18/2024 06:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.3 F 0.0
2:00 161.1 F 3.2
4:00 189.6 F 8.3
6:00 213.7 F 8.4
8:00 231.1 F 9.4
10:00 234.6 F 9.0
12:00 244.5 F 13.2
14:00 243.8 F 12.3
16:00 257.3 F 19.6
18:00 269.1 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.7
1:00 272.8 F 29.2
2:00 272.5 F 29.5
3:00 271.8 F 28.6
4:00 272.7 F 29.2

Min 271.8 F 28.6
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 218.3 F 0.0
10:00 200.9 F 0.0
15:00 203.8 F 0.0
20:00 201.9 F 0.0
25:00 199.7 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 18:50
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

09/18/2024 07:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____