

-----  
Midmark M11 - v1.0.5  
Total Cycles: 411

Name: RITTER2  
Sterilizer ID: V2313228  
09/30/2021 11:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 123.5 F 0.0  
2:00 130.4 F 1.8  
4:00 161.7 F 5.5  
6:00 186.3 F 10.2  
8:00 209.6 F 4.5  
10:00 225.2 F 9.0  
12:00 233.1 F 10.5  
14:00 239.2 F 11.5  
16:00 238.1 F 10.0  
18:00 251.6 F 16.6  
20:00 263.2 F 23.4  
22:00 272.0 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.0  
1:00 272.4 F 29.1  
2:00 272.7 F 29.5  
3:00 272.0 F 29.6  
4:00 272.3 F 29.0

Min 271.6 F 28.9  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.6 F 0.0  
5:00 212.0 F 0.0  
10:00 203.3 F 0.0  
15:00 205.4 F 0.0  
20:00 208.4 F 0.0  
25:00 203.2 F 0.0  
30:00 200.6 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 22:05  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:00:37  
CYCLE COMPLETE

09/30/2021 12:42 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02