

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3605

Name: RITTER2  
Sterilizer ID: V2313228  
05/10/2024 09:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 127.0 F 0.0  
2:00 128.0 F 1.1  
4:00 135.6 F 3.4  
6:00 160.9 F 6.6  
8:00 183.8 F 10.4  
10:00 208.5 F 6.0  
12:00 218.6 F 7.4  
14:00 230.4 F 11.7  
16:00 230.1 F 8.0  
18:00 241.2 F 13.3  
20:00 241.2 F 11.8  
22:00 252.7 F 17.6  
24:00 263.0 F 23.8  
26:00 272.8 F 30.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.5  
1:00 272.8 F 30.3  
2:00 272.9 F 30.2  
3:00 272.2 F 30.5  
4:00 272.1 F 29.9

Min 270.9 F 29.3  
Max 273.5 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.4 F 0.0  
5:00 218.6 F 0.0  
10:00 199.8 F 0.0  
15:00 196.7 F 0.0  
20:00 211.0 F 0.0  
25:00 218.0 F 0.0  
30:00 219.1 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 26:27  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:04:59  
CYCLE COMPLETE

05/10/2024 10:59 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_