
Midmark M11 - v1.0.5
Total Cycles: 3666

Name: RITTER2
Sterilizer ID: V2313228
06/04/2024 11:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.4 F 0.0
2:00 99.1 F 0.3
4:00 111.7 F 1.5
6:00 130.4 F 3.7
8:00 166.2 F 6.8
10:00 182.6 F 10.3
12:00 209.4 F 7.1
14:00 219.3 F 7.0
16:00 225.5 F 10.9
18:00 229.5 F 7.1
20:00 240.5 F 11.6
22:00 237.2 F 9.4
24:00 249.3 F 15.2
26:00 259.3 F 21.0
28:00 268.4 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 29.2
1:00 271.6 F 29.2
2:00 271.9 F 29.6
3:00 272.0 F 29.5
4:00 272.5 F 30.5

Min 271.0 F 29.1
Max 272.7 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.1 F 0.0
5:00 226.5 F 0.0
10:00 198.3 F 0.0
15:00 176.9 F 0.0
20:00 174.5 F 0.0
25:00 172.5 F 0.0
30:00 171.1 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 28:55
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:07:22
CYCLE COMPLETE

06/04/2024 12:34 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____