
Midmark M11 - v1.0.5
Total Cycles: 291

Name: RITTER2
Sterilizer ID: V2313228
08/19/2021 11:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.3 F 0.0
2:00 177.6 F 4.2
4:00 198.6 F 9.4
6:00 222.7 F 6.9
8:00 231.2 F 10.7
10:00 234.0 F 8.6
12:00 240.2 F 10.2
14:00 249.7 F 15.4
16:00 262.8 F 22.9
18:00 271.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.3 F 28.9
2:00 272.7 F 29.5
3:00 271.7 F 28.9
4:00 272.6 F 29.2

Min 271.6 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.8 F 0.0
5:00 234.0 F 0.0
10:00 225.0 F 0.0
15:00 215.0 F 0.0
20:00 229.7 F 0.0
Dry Cycle Aborted

FILLING: 2:51
HEATING: 18:01
STERILIZING: 4:00
VENTING: 3:05
DRYING: 23:35
TOTAL CYCLE: 00:51:32

08/19/2021 12:44 PM
ACCEPTED: X
REJECTED:
Operator: U02