

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1797

Name: RITTER2  
Sterilizer ID: V2313228  
12/02/2022 10:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.5 F 0.0  
2:00 79.6 F 0.0  
4:00 108.7 F 0.0  
6:00 121.8 F 0.1  
8:00 168.0 F 0.3  
10:00 197.8 F 3.2  
12:00 211.5 F 7.7  
14:00 224.7 F 12.2  
16:00 229.0 F 7.4  
18:00 240.7 F 12.5  
20:00 239.6 F 10.3  
22:00 237.1 F 9.0  
24:00 252.5 F 16.5  
26:00 264.9 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.4  
1:00 272.0 F 28.7  
2:00 272.4 F 29.1  
3:00 272.8 F 29.2  
4:00 272.9 F 29.1

Min 271.7 F 28.2  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 233.0 F 0.0  
5:00 214.6 F 0.0  
10:00 223.7 F 0.0  
15:00 226.0 F 0.0  
20:00 223.3 F 0.0  
25:00 221.6 F 0.0  
30:00 219.6 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 27:36  
STERILIZING: 4:00  
VENTING: 2:40  
DRYING: 30:00  
TOTAL CYCLE: 01:05:04  
CYCLE COMPLETE

12/02/2022 11:06 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_