
Midmark M11 - v1.0.5
Total Cycles: 1438

Name: RITTER2
Sterilizer ID: V2313228
08/31/2022 10:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.4 F 0.0
2:00 133.3 F 0.0
4:00 149.6 F 0.8
6:00 184.0 F 3.3
8:00 206.6 F 6.9
10:00 220.7 F 11.5
12:00 227.7 F 6.9
14:00 238.3 F 12.0
16:00 236.8 F 9.6
18:00 237.3 F 9.4
20:00 248.2 F 14.7
22:00 260.0 F 21.3
24:00 270.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.0
1:00 272.4 F 29.7
2:00 271.6 F 28.9
3:00 272.3 F 29.6
4:00 271.6 F 29.0

Min 271.4 F 28.9
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 211.2 F 0.0
10:00 223.4 F 0.0
15:00 222.3 F 0.0
20:00 221.0 F 0.0
25:00 221.1 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:44
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 01:02:36
CYCLE COMPLETE

08/31/2022 11:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____