

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5477

Name: RITTER 1  
Sterilizer ID: V2178198  
04/04/2025 08:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.8 F 0.0  
2:00 74.6 F 0.4  
4:00 91.1 F 1.8  
6:00 124.3 F 4.1  
8:00 148.8 F 6.9  
10:00 169.1 F 10.2  
12:00 195.7 F 9.4  
14:00 204.8 F 3.7  
16:00 216.5 F 6.9  
18:00 226.6 F 10.5  
20:00 228.4 F 8.0  
22:00 237.6 F 10.0  
24:00 236.9 F 8.8  
26:00 246.1 F 13.2  
28:00 257.5 F 19.2  
30:00 267.8 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.5 F 28.7  
2:00 272.1 F 28.6  
3:00 272.1 F 28.4  
4:00 272.6 F 29.0

Min 271.6 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.3 F 0.0  
5:00 182.5 F 0.0  
10:00 191.4 F 0.0  
15:00 198.6 F 0.0  
20:00 201.6 F 0.0  
25:00 202.3 F 0.0  
30:00 202.8 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:23  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:09:38  
CYCLE COMPLETE

04/04/2025 09:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_