
Midmark M11 - v1.0.5
Total Cycles: 401

Name: RITTER2
Sterilizer ID: V2313228
09/28/2021 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.5 F 0.0
2:00 147.9 F 2.2
4:00 164.7 F 5.8
6:00 188.3 F 10.5
8:00 214.5 F 4.8
10:00 224.9 F 10.3
12:00 229.9 F 7.7
14:00 243.0 F 13.1
16:00 242.6 F 11.9
18:00 255.2 F 18.6
20:00 266.8 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.3 F 29.0
2:00 272.0 F 29.5
3:00 271.6 F 29.4
4:00 272.1 F 29.1

Min 271.2 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 232.0 F 0.0
10:00 207.9 F 0.0
15:00 222.6 F 0.0
20:00 165.7 F 0.0
25:00 153.7 F 0.0
30:00 140.9 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:26
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:59:39
CYCLE COMPLETE

09/28/2021 12:55 PM
ACCEPTED: X
REJECTED:
Operator: U01