
Midmark M11 - v1.0.0
Total Cycles: 1173

Name: RITTER 1
Sterilizer ID: V2178198
09/15/2021 06:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.1 F 0.0
2:00 78.5 F 0.0
4:00 88.8 F 0.0
6:00 116.7 F 0.1
8:00 140.6 F 0.1
10:00 169.2 F 0.2
12:00 192.4 F 2.4
14:00 201.7 F 5.2
16:00 212.1 F 8.3
18:00 223.6 F 11.8
20:00 226.2 F 6.4
22:00 237.2 F 11.2
24:00 234.4 F 8.2
26:00 243.5 F 11.6
28:00 245.7 F 13.0
30:00 257.3 F 19.2
32:00 267.8 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.7 F 29.0
2:00 271.8 F 28.2
3:00 272.1 F 28.7
4:00 272.5 F 28.8

Min 271.5 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.4 F 0.0
5:00 181.3 F 0.0
10:00 179.3 F 0.0
15:00 181.9 F 0.0
20:00 180.2 F 0.0
25:00 170.0 F 0.0
30:00 169.7 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 33:25
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:11:37
CYCLE COMPLETE

09/15/2021 07:46 AM
ACCEPTED: X
REJECTED:
Operator: U01