

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4860

Name: RITTER 1  
Sterilizer ID: V2178198  
08/28/2024 02:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 87.8 F 0.0  
2:00 88.3 F 0.0  
4:00 109.4 F 0.5  
6:00 137.7 F 0.9  
8:00 166.2 F 1.4  
10:00 189.3 F 4.6  
12:00 211.3 F 8.3  
14:00 220.9 F 12.4  
16:00 224.7 F 6.5  
18:00 235.3 F 11.3  
20:00 233.7 F 8.1  
22:00 243.8 F 11.8  
24:00 244.6 F 12.4  
26:00 256.3 F 18.5  
28:00 266.6 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.4 F 28.9  
2:00 271.4 F 28.1  
3:00 272.3 F 29.0  
4:00 271.9 F 28.5

Min 271.1 F 28.0  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.6 F 0.0  
5:00 205.8 F 0.0  
10:00 206.5 F 0.0  
15:00 200.5 F 0.0  
20:00 201.9 F 0.0  
25:00 214.9 F 0.0  
30:00 214.7 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 29:40  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:08:02  
CYCLE COMPLETE

08/28/2024 03:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_