
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001949
7:20 14/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:57
STERILIZING 4:00
136.3 C 222kPa 4:00
136.4 C 225kPa 4:31
136.4 C 224kPa 5:01
136.4 C 224kPa 5:31
136.3 C 221kPa 6:01
136.4 C 225kPa 6:31
136.2 C 222kPa 7:01
136.4 C 221kPa 7:31
136.4 C 221kPa 7:31
VENTING 7:31
AIR DRYING 8:21
CYCLE COMPLETE 8:25

Digital Signature #
8A886199A28EEFBF
