

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000718  
9:53 21/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:12  
STERILIZING 4:20  
136.0 C 228kPa 4:20  
136.2 C 232kPa 4:50  
136.4 C 232kPa 5:20  
136.4 C 232kPa 5:50  
136.2 C 232kPa 6:20  
136.4 C 233kPa 6:50  
136.4 C 233kPa 7:20  
136.4 C 233kPa 7:50  
136.4 C 233kPa 7:50  
VENTING 7:50  
AIR DRYING 8:47  
CYCLE COMPLETE 8:48  
-----

Digital Signature #  
0CDC05C07280850F  
-----