

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000480  
15:43 26/02/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:13  
STERILIZING 4:23  
136.0 C 228kPa 4:23  
136.4 C 233kPa 4:53  
136.3 C 231kPa 5:23  
136.4 C 233kPa 5:53  
136.4 C 233kPa 6:23  
136.3 C 232kPa 6:53  
136.2 C 231kPa 7:23  
136.4 C 233kPa 7:53  
136.4 C 233kPa 7:53  
VENTING 7:53  
AIR DRYING 8:50  
CYCLE COMPLETE 15:14  
-----

Digital Signature #  
4650A39EBAF5EE95  
-----