
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000093
14:58 28/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:13
STERILIZING 4:28
136.0 C 220kPa 4:28
136.4 C 222kPa 4:58
136.1 C 222kPa 5:28
136.2 C 224kPa 5:58
136.4 C 226kPa 6:28
136.2 C 224kPa 6:58
136.3 C 225kPa 7:28
136.1 C 221kPa 7:58
136.1 C 221kPa 7:58
VENTING 7:58
AIR DRYING 8:50
CYCLE COMPLETE 10:38

Digital Signature #
B3BAD5E05CA8A5AB
