

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1646

Name: RITTER2  
Sterilizer ID: V2313228  
10/24/2022 01:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.1 F 0.0  
2:00 188.4 F 3.9  
4:00 203.0 F 9.0  
6:00 222.5 F 7.1  
8:00 230.3 F 10.2  
10:00 234.6 F 8.1  
12:00 242.9 F 12.3  
14:00 244.8 F 13.1  
16:00 257.4 F 19.7  
18:00 268.4 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.4 F 29.5  
2:00 271.4 F 28.8  
3:00 272.3 F 29.7  
4:00 272.0 F 29.2  
  
Min 271.2 F 28.8  
Max 272.5 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.8 F 0.0  
5:00 219.4 F 0.0  
10:00 237.6 F 0.0  
15:00 239.0 F 0.0  
20:00 236.9 F 0.0  
25:00 227.5 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 2:07  
HEATING: 19:14  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:29  
CYCLE COMPLETE

10/24/2022 02:42 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_