

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2475

Name: RITTER 1  
Sterilizer ID: V2178198  
10/24/2022 09:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.8 F 0.0  
2:00 171.8 F 3.1  
4:00 178.1 F 7.4  
6:00 200.9 F 12.4  
8:00 213.8 F 5.7  
10:00 233.1 F 10.8  
12:00 230.9 F 7.1  
14:00 244.5 F 13.3  
16:00 243.1 F 11.7  
18:00 255.5 F 18.3  
20:00 266.8 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.6 F 29.0  
2:00 271.8 F 28.4  
3:00 271.7 F 28.7  
4:00 272.5 F 29.0  
  
Min 271.3 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.4 F 0.0  
5:00 236.0 F 0.0  
10:00 231.6 F 0.0  
15:00 230.1 F 0.0  
20:00 230.2 F 0.0  
25:00 227.2 F 0.0  
30:00 229.1 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 21:32  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:59:43  
CYCLE COMPLETE

10/24/2022 09:59 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_